



We Own the Future of Bolting Technology

Certified Report

<u>Customer:</u>	<u>Address:</u>	<u>Job No.:</u>	<u>Report No.:</u>
Pressure Reading (psi)	Pressure Reading (bar)	Torque Reading (ft.-lb.)	Torque Reading (N-m)
1000	69	149	202
1200	83	178	241
1400	97	208	282
1600	110	238	323
1800	124	267	362
2000	138	297	403
2200	152	327	443
2400	166	356	483
2600	179	386	523
2800	193	416	564
3000	207	446	605
3200	221	475	644
3400	234	505	685
3600	248	535	725
3800	262	564	765
4000	276	594	805
4200	290	624	846
4400	303	653	885
4600	317	683	926
4800	331	716	971
5000	345	743	1008
5200	359	772	1047
5400	372	802	1088
5600	386	832	1128
5800	400	861	1168
6000	414	891	1208
6200	428	921	1249
6400	441	950	1288
6600	455	980	1329
6800	469	1010	1370
7000	483	1040	1410
7200	497	1069	1450
7400	510	1099	1490
7600	524	1129	1531
7800	538	1158	1570
8000	552	1188	1611
8200	566	1218	1652
8400	579	1247	1691
8600	593	1277	1732
8800	607	1306	1771
9000	621	1334	1809
9200	634	1366	1852
9400	648	1395	1892
9600	662	1426	1934
9800	676	1455	1973
10000	690	1486	2015

*** CERTIFICATION ***

This instrument has been tested to comply with the relevant specifications in accordance with the above purchase order. The transducer used to obtain the above values is traceable to N.I.S.T. Standards.

<u>Manufacturer</u>	<u>Model</u>	<u>Serial No.</u>	<u>Customer I.D.</u>
Boltech	QD2SQDR		
<u>Date:</u>	<u>Signature of Technician Performing Calibration</u>		<u>Calibrated Transducer S/N:</u>
			35361.LOG